

CERTIFICATE OF CALIBRATION

Instruments details

Model number: PH818
Description: pH Meter
Serial Number: 6233027
Date of manufacture: 2022-08-12

Reference document for the calibration:

JJG 919-1996

Place and environmental conditions of the calibration:

Place	Temperature	RH	
SMART SENSOR FQC	25 ±2°C	55% ±5%RH	
Function Range Mode	Normalized Value	Actual Value	Permissible Error
	(pH)	(pH)	(pH)
Measure Value	4.01	4.02	0.01
	6.86	6.85	-0.01
	9.18	9.17	-0.01

