



HI3887

## Quick-Check Swimming Pool Test Kit

*Free Chlorine and pH*


The HI3887 is a colorimetric chemical test kit that determines the free chlorine concentration and pH level in samples within a 0.0 to 2.5 mg/L (ppm)  $\text{Cl}^-$  range and 6.0 to 8.5 pH range. The HI3887 is supplied with all of the necessary reagents and equipment to perform the analysis. The test kit contains enough reagents to perform approximately 50 tests for free chlorine and 100 tests for pH.

- **Complete setup**
  - All required materials are included with the test kit, such as the color comparison cubes and reagent dropper bottles.
- **High resolution**
  - Free chlorine readings from 0.0 to 2.5 mg/L are determined to 0.5 mg/L resolution.
  - pH readings from 6.0 to 8.5 pH are determined to 0.5 pH resolution.

### Significance of Use

Chlorine is one of the most commonly used disinfectants for drinking water, wastewater, and water used for pools and spas. It can be added to in various forms including calcium hypochlorite, sodium

hypochlorite, or in some instances, chlorine gas. When added to water, chlorine creates hypochlorous acid ( $\text{HOCl}$ ) which dissociates into hypochlorite ion ( $\text{OCl}^-$ ).



hypochlorous acid  $\leftrightarrow$  hydrogen ion + hypochlorite ion

$\text{HOCl}$  is the form of chlorine that acts as a stronger disinfectant as compared to  $\text{OCl}^-$ . To ensure the added chlorine is effective at sanitizing, the pH of the water must be taken into account. Around pH 7.5,  $\text{HOCl}$  and  $\text{OCl}^-$  are present in relatively equal amounts. Below pH 7.5, the equilibrium shifts to favor  $\text{HOCl}$ ; above pH 7.5, the equilibrium shifts to favor  $\text{OCl}^-$ . Depending on the application, addition of chlorine is effective when added to water with a neutral or slightly acidic pH value.

When chlorine is first added to water, it is available as free chlorine. The measurement of free chlorine signifies the amount available for disinfection. Once chlorine begins to sanitize bacteria and pathogens present in the water, it becomes combined chlorine; combined chlorine is no longer available to act as a disinfectant.

Specifications	HI3887 Quick-Check Swimming Pool Test Kit				
Test	Type	Range	Smallest Increment	Method	Number of Tests
Free Chlorine	colorimetric	0-2.5 mg/L (ppm)	0.5 mg/L (ppm)	DPD	50 avg.
pH	colorimetric	6.0-8.5 pH	0.5 pH	pH indicator	100 avg.
Ordering Information	<b>HI3887</b> test kit includes color comparison cubes (2), 20 mL reagent 1, 12 mL reagent 2, 25 mL pH reagent and instructions.				
Reagents	<b>HI3831F-050</b> free chlorine, 50 tests avg.				

ติดต่อบริษัท เนโอเนิกส์ จำกัด

Tel: 098-479-5684 หรือ 061-8268939

E-mail: sale@neonics.co.th หรือ sale@tools.in.th